K973238

JAN 1 4 2000

510(k) SUMMARY

SUMMARY OF SAFETY AND EFFECTIVENESS

Fortoss-CemaTM

In accordance with the requirements of the Safe Medical Devices Act of 1990 and 21 CFR 807.92 Biocomposites Ltd is hereby submitting the 510(k) Summary of Safety and Effectiveness for Fortoss-CemaTM.

A. Submitter

Biocomposites Ltd Etruscan Street Etruria Stoke-on-Trent England ST1 5PQ

B. Contact Person

J Stephen Bratt LL.B

C. Date Prepared

22nd September 1999

D. Name of the Device

Fortoss-Cema™

E. Common or Usual Name

Calcium Sulfate Hemihydrate

F. Classification Name

Implant, endosseous for bone filling and/or augmentation

G. Predicate/Legally Marketed Devices

Lifecore Biomedical CapSet™

H. Device Description

Fortoss-Cema™ is a medical grade calcium sulfate powder and mixing diluent. The blended materials form a creamy paste which set to form a barrier over bone graft materials.

I. Intended Use

Fortoss-CemaTM is used as a barrier over bone graft materials to prevent particle migration in an osseous defect.

J. Substantial Equivalence

Fortoss-CemaTM is substantially equivalent in design, function and intended use to the predicate device Lifecore Biomedical CapSetTM.



Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

JAN 1 4 2000

Mr. J.S. Bratt Managing Director Biocomposites, Ltd. Etruscan Street, Etruria, Stroke-on-Trent Staffordshire, ST1, 5PQ, England

Re: K993238

Trade Name: Fortoss CemaTM Calcium Sulfate Barrier

Regulatory Class: Unclassified

Product Code: LYC

Dated: December 17, 1999 Received: December 20, 1999

Dear Mr. Bratt:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the <u>Code of Federal Regulations</u>, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Good Manufacturing Practice for Medical Devices: General (GMP) regulation (21 CFR Part 820) and that, through periodic GMP inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any

obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4692. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "http://www.fda.gov/cdrh/dsmamain.html".

Sincerely yours,

Timothy A. Ulatowski

Director

Division of Dental, Infection Control and General Hospital Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

K993238

Page 1 of 1

510(k) Number (if known):	K993238
Device Name:	Fortoss Cema TM
Indications For Use:	
Fortoss-Cema TM is a resorba	ble pre-measured powder which, when blended with the
supplied sterile diluent, forms a	a setting paste.
Fortoss-Cema TM is indicated for use as a barrier in placement over bone graft material to provide a resorbable barrier and prevent particle migration in an osseous defect.	
(PLEASE DO NOT WRITE BE IF NEEDED)	ELOW THIS LINE-CONTINUE ON ANOTHER PAGE
Concurrence of CDRH, Office of Device Evaluation (ODE)	
Concurrence of objects in a second of the se	
Prescription Use	OR Over-The-Counter Use
(Per 21 CFR 801.109)	(Optional Format 1-2-
96)	\cap
(Divis	ion Sign-Off)
Divisio	on of Dental, Infection Control, eneral Hospital Devices
	wordt riospital Devices

K99 3238

510(k) Number____